



Series SAH SMART Air Hood™ Balancing Instrument

Recalibration Return Pack

Thank you for purchasing the Dwyer SMART Air Hood™ Balancing Instrument. For your convenience, Dwyer has included this set of 5 boxes for the purpose of returning the four sensing grids and the sensor module for recalibration. The pack also includes a self-addressed shipping label.

Dwyer recommends recalibrating the hood, at a minimum, once a year to ensure optimum performance. Dwyer offers two options for recalibrating the Series SAH.

Option 1: To ensure no downtime, you can purchase a new set of four calibrated sensing grids and sensor module, part number 100765-09, which are calibrated together and ready to install. Pricing for this option is available at www.dwyer-inst.com.

Option 2: Return the original four sensing grids and sensor module in the recalibration return pack to Dwyer using our RMA process or to any authorized Dwyer calibration service company. Once the sensing grids and sensor module are recalibrated, they will be returned as a set back to you, ready to install. Pricing for Dwyer recalibration services is available at www.dwyer-inst.com. If returning to Dwyer directly for recalibration, contact Customer Service for an RMA number.

To remove the sensing grids, simply detach by lifting up on the two locking tabs on each side of the grids. The sensor module is removed by pushing in on the two tabs on the ends simultaneously and lifting. Be sure to return all four of the sensing grids and the sensor module as they need to be calibrated together.

