

# Manual Supplement

Manual Title:	753/754 Calibration	Supplement Issue:	<b>2</b>
Part Number:	Web Only	Issue Date:	12/12
Print Date:	November 2011	Page Count:	1
Revision/Date:			

---

---

This supplement contains information necessary to ensure the accuracy of the above manual. Enter the corrections in the manual if either one of the following conditions exist:

1. The revision letter stamped on the indicated PCA is equal to or higher than that given with each change.
2. No revision letter is indicated at the beginning of the change.

## Change #1, 61489

On page 8, in the **Temperature, Thermocouples** Table:

Change: Normal Mode Rejection: 40 dB at 50 Hz or 60 Hz nominal

To: Normal Mode Rejection: 65 dB at 50 Hz or 60 Hz nominal

## Change #2, 64190, 64295

On page 3, add the following to the **Symbols** Table:

	Conforms to CAN/CSA-C22.2 No. 61010-1, second edition, including Amendment 1.
	Conforms to relevant South Korean EMC Standards.

On page 4, replace the **Standard and Agency Approval Specification** with:

### **Standards and Agency Approval Specifications**

<b>Intrusion Protection</b> .....	IEC 60529-1, IP 52
<b>Safety</b> .....	IEC 61010-1, CAT II 300 V, Pollution degree 2
<b>Electromagnetic Environment</b> .....	IEC 61326-1, Basic
<b>Electromagnetic Compatibility</b> .....	Accuracy for all function is not specified in RF fields >3 V/m

Applies to use in Korea only. Class A Equipment (Industrial Broadcasting & Communication Equipment) <sup>[1]</sup>

[1] This product meets requirements for industrial (Class A) electromagnetic wave equipment and the seller or user should take notice of it. This equipment is intended for use in business environments and is not to be used in homes.