

- $10m\Omega$ to $12k\Omega$
- 0.01% accuracy
- Precision PT100 simulation
- Low temperature coefficient
- In-line readout
- Excellent long term stability
- 6 digit resolution
- Fully screened

DESCRIPTION

A precision resistance decade box suitable for a wide range of simulation work. High accuracy, long term stability, and low temperature coefficient make the 1067 ideal for simulating and calibrating precision PT100 sensors and temperature indicators/meters that use resistive sensors.

Special care has been taken in the construction of the 1067 to ensure that the residual end resistance is as low and as stable as possible. Multiple self-cleaning silver alloy contacts are used for each position to ensure outstanding performance and long life.

Housed in a robust metal case the 1067 is fully screened and low thermal emf terminals are used. The slim line design means it takes up minimum bench space and is easily transportable.

Resistance is selected by dialling the value required using the rotary switches. This enables precise setting with a clear unambiguous indication. Each decade is scaled from 0 to 11 and therefore allows convenient overlap of the set values. The maximum value settable is 12,222.21 ohms.

SPECIFICATIONS

0 to 12k Ω / 10m Ω steps
\pm 0.01% of setting \pm 2m Ω , after deduction of residual end resistance \pm 1m Ω for residual variation. At calibration temperature of 22°C.
Less than 10m Ω . Less than 1m Ω variation
0.35 watt per resistor
Maximum 200V DC/AC RMS
20ppm/year (>1 Ω), 100ppm/year (<1 Ω)
Less than 10ppm/°C (> 1Ω). Less than 20ppm/°C (< 1Ω)
Less than 0.1Nm
Precision Resistance Decade Box

Time Electronics Ltd, Unit 11 Sovereign Way, Botany Industrial Estate, Tonbridge, Kent, TN9 1RH. United Kingdom. T: +44 (0) 1732 355993 F: +44 (0) 1732 770312 E: mail@timeelectronics.co.uk

C161 Factory Calibration Certificate (NPL)
C114 UKAS Calibration Certificate (ISO 17025)