

Instruction Sheet

PN 51A-371pH/rev.D

January 2011

EuroSenz™ Model 371 pH/ORP Sensor and Model 370 pH Sensor

Low Maintenance pH Combination Electrode

For additional information please visit our website at
www.emersonprocess.com/raihome/liquid/.

SPECIFICATIONS

pH Range: 0-14 pH

ORP Range: ± 1500 mV

Glass Type: General Purpose Low Resistance (GPLR)

Percent Linearity: 96% @ 0-4 pH, 99% @ 4-12 pH,
97% @ 12-13 pH, 92% @ 13-14 pH

Temperature Range: 0° - 85°C (32° - 185°F)

Pressure Range: 100-790 kPa abs (0-100 psig)

Conductivity: >100mS/cm

Wetted Materials: Polypropylene, EP, and Glass

Process Connection: Standard PG 13.5 Thread

Cable Connection: Eurocab (-70); VP (-72)

S8 Cable: Single Pole Euro PN:9200338

VP8 Cable: PN: 24281-00



370 and 371-70 Eurocap



371-72 (VP sensor)

⚠ WARNING



Before removing the sensor, be absolutely certain that the process pressure is reduced to 0 psig and the process temperature is lowered to a safe level!

⚠ CAUTION

SENSOR/PROCESS APPLICATION COMPATIBILITY

The wetted sensor materials may not be compatible with process composition and operating conditions. Application compatibility is entirely the responsibility of the user.

ATEX DIRECTIVE

Special Conditions for safe use

1. All pH/ORP sensors have a plastic enclosure which must only be cleaned with a damp cloth to avoid the danger due to a build up of an electrostatic charge.
2. All pH/ORP sensor Models are intended to be in contact with the process fluid and may not meet the 500V r.m.s. a.c. test to earth. This must be taken into consideration at installation.

STORAGE

1. It is recommended that electrodes be stored in their original shipping containers until needed.
2. Do not store at temperatures below -5°C (23°F).
3. Electrodes should be stored with a protective cap containing KCl solution (PN 9210342).
4. For overnight storage, immerse the sensor in tap water or 4 pH buffer solution.
5. A pH glass electrode has a limited shelf life of one year.

ELECTRODE PREPARATION

1. Remove electrode from shipping container.
2. Remove the protective boot covering the electrode bulb.
3. Rinse away salt film with clean water, then shake the electrode so that the internal solution fills the bulb, thus removing any air trapped there.



NOTE

Do not allow lubricant to coat electrode bulb or reference junction. If it does, wipe it clean before installation.

INSTALLATION

For sensor orientation, see Figure 1.

For wiring, see Figures 3 and 4.



TWO POINT BUFFER CALIBRATION

Select two stable buffer solutions, preferably pH 4.0 and 10.0 (pH buffers other than pH 4.0 and pH 10.0 can be used as long as the pH values are at least two pH units apart).

NOTE

A pH 7 buffer solution reads a mV value of approx. zero, and pH buffers read approximately ± 59.1 mV for each pH unit above or below pH 7. Check the pH buffer manufacturer specifications for millivolt values at various temperatures since it may affect the actual value of the buffer solution mV/pH value.

1. Immerse sensor in the first buffer solution. Allow sensor to equilibrate to the buffer temperature (to avoid errors due to temperature differences between the buffer solution and sensor temperature) and wait for reading to stabilize. Value of buffer can now be acknowledged by analyzer/transmitter.
2. Once the first buffer has been acknowledged by the analyzer/transmitter, rinse the buffer solution off of the sensor with distilled or deionized water.
3. Repeat steps 1 and 2 using the second buffer solution.
4. The theoretical slope value, according to the Nernst equation for calculating pH, is approximately 59.17 mV/pH. Over time the sensor will age, both in the process and in storage, and will result in reduced slope values. To ensure accurate readings, it is recommended that the electrode be replaced when the slope value falls below 47 to 49 mV/pH.

RECOMMENDED pH SENSOR STANDARDIZATION





For maximum accuracy, the sensor can be standardized on-line or with a process grab sample after a buffer calibration has been performed and the sensor has been conditioned to the process. Standardization accounts for the sensor junction potential and other interferences. Standardization will not change the sensor's slope but will simply adjust the analyzer's reading to match that of the known process pH.

MAINTENANCE

Electrodes should respond rapidly. Sluggishness, offsets, and erratic readings are indicators that the electrodes may need cleaning or replacement.

1. To remove oil deposit, clean the electrode with a mild non-abrasive detergent.
2. To remove scale deposits, soak electrodes for 30 to 60 minutes in a 5% hydrochloric acid solution.
3. Temperature effect on life expectancy: If glass electrode life expectancy is 100% @ 25°C (77°F), then it will be approximately 25% @ 80°C (176°F), and approximately 5% @ 120°C (248°F).

WIRING






FIGURE 9. VP8 Wiring to Xmt analyzer. Model 371-72




FIGURE 10. 371-72 wiring to Model 1055-01-10-22 single sensor wiring VP6 (28645-XX)




FIGURE 11. 371-72 Wiring to Model 1055-01-10-22-32 dual sensor wiring VP6 (28645-XX)


ACCESSORIES

PN	DESCRIPTION
9200338	Cable with single pole connector for Eurocap (option -70 and Model 370), coaxial connection, 5 m (16.4 ft)
9200339	Cable with single pole connector for Eurocap (option -70 and Model 370), coaxial connection, 10 m (33 ft)
9120550	Cable with VP 6.0 connector (option -72), coaxial connection with TC (option -72), 3 m (9.8 ft)
9160504	Cable with VP 6.0 connector (option -72), coaxial connection with TC (option -72), 5 m (16.4 ft)
9210012	Buffer solution, 4.01 pH, 500 ml (16 oz)
9210013	Buffer solution, 6.86 pH, 500 ml (16 oz)
9210014	Buffer solution, 9.18 pH, 500 ml (16 oz)
9120342	KCl electrolyte solution, 250 ml (8 oz)
R508-802	ORP standard solution, 460 mV, 500 ml (8 oz)
24281-00	15 ft. cable with mating VP8 connector; 371 sensor option (-72)
24281-01	25 ft. cable with mating VP8 connector; 371 sensor option (-72)
24281-02	2.5 ft. cable with mating VP8 connector; 371 sensor option (-72)
24281-03	50 ft. cable with mating VP8 connector; 371 sensor option (-72)
24281-04	100 ft. cable with mating VP8 connector; 371 sensor option (-72)
24281-05	4 ft. cable with mating VP8 connector; 371 sensor option (-72)
24281-06	10 ft. cable with mating VP8 connector; 371 sensor option (-72)
24281-07	20 ft. cable with mating VP8 connector; 371 sensor option (-72)
24281-08	30 ft. cable with mating VP8 connector; 371 sensor option (-72)



*The right people,
the right answers,
right now.*

ROSEMOUNT ANALYTICAL
CUSTOMER SUPPORT CENTER
1-800-854-8257



ON-LINE ORDERING NOW AVAILABLE ON OUR WEB SITE
<http://www.raihome.com>

Specifications subject to change without notice.



Credit Cards for U.S. Purchases Only.



Emerson Process Management

2400 Barranca Parkway
Irvine, CA 92606 USA
Tel: (949) 757-8500
Fax: (949) 474-7250

<http://www.raihome.com>

© Rosemount Analytical Inc. 2011

