

Rosemount 751 Field Signal Indicator Operating Temperature Limit Change

Dear valued Emerson customer,

The Rosemount 751 Field Signal Indicator with the LCD display operating temperature limits are: -40 °C to 70 °C (-40 °F to 158 °F). Due to recent testing, the updated operating temperature limits will remain unchanged at -40 °C to 70 °C (-40 °F to 158 °F). However, for temperatures below -20 °C (-4 °F) or above 60 °C (140 °F), the LCD display may not be readable, but the loop will remain intact and the LCD display will not be damaged. In most circumstances, the LCD display will be readable over 60 °C (140 °F); however, if the LCD display is installed in an application where the loop current is below 6.0 mA, the LCD display may become unreadable.

We at Emerson Process Management are committed to supplying our customers with the best measurement products. Thank you for allowing Emerson to serve you now and in the future. Please contact your local Emerson Process Management representative for additional concerns.

*The Emerson logo is a trademark and service mark of Emerson Electric Co.
Rosemount and the Rosemount logotype are registered trademarks of Rosemount Inc.
All other marks are the property of their respective owners.*

© 2012 Rosemount Inc. All rights reserved.

Emerson Process Management

Rosemount Inc.
8200 Market Boulevard
Chanhassen, MN 55317 USA
T (U.S.) 1-800-999-9307
T (International) (952) 906-8888
F (952) 906 8889

ROSEMOUNT®

www.rosemount.com


EMERSON™
Process Management